

2	0.947	-5.30%	0.011	1.16%	Passed
3	0.943	-5.70%	0.007	0.74%	Passed
4	0.948	-5.20%	0.006	0.63%	Passed
5	0.938	-6.20%	0.024	2.56%	Passed
6	0.944	-5.60%	0.010	1.06%	Passed
7	0.935	-6.50%	0.010	1.07%	Passed
8	0.926	-7.40%	0.010	1.08%	Passed

Group 1 Dispense Order Statistics:

Dispense	Mean Volume	Inaccuracy	Standard Deviation	CV
1	0.933	-6.70%	0.012	1.29%
2	0.934	-6.60%	0.011	1.18%
3	0.934	-6.60%	0.008	0.86%
4	0.941	-5.90%	0.016	1.70%
5	0.943	-5.70%	0.015	1.59%
6	0.942	-5.80%	0.010	1.06%
7	0.935	-6.50%	0.030	3.21%
8	0.945	-5.50%	0.011	1.16%
9	0.945	-5.50%	0.006	0.63%
10	0.941	-5.90%	0.007	0.74%
11	0.940	-6.00%	0.007	0.74%
12	0.946	-5.40%	0.009	0.95%

Plate Reader:

Reader Type	Serial Number	Calibrated
Artel ELx800NB®		4/17/2012 9:53:39 AM

Materials:

Item Description	Lot or Serial No.	Expiration Date
Calibrator Plate	1195	5/11/2012
Baseline Solution	Z11208111501	3/8/2013
MVS Verification Plate - Baseline	13211022-00469	10/28/2016
Range D Solution	RD051611150122062	8/16/2012
Diluent Solution	U11122111501	2/22/2013
MVS Verification Plate	13211022-00467	10/28/2016

Performed By:

User Name	User ID	Date	Time	E-Signature
Megan Mathieson	mlhf5	4/17/2012	11:38:17 AM GMT10	c4bae9f69ee925f37e30d08bbbbaad7c4